



Type examination certificate

No. CH-MI001-11052-00

Bauartprüfzertifikat

Applicant:
Auftraggeber:

G.GIOANOLA SRL
Strada Alessandria 50
14049 Nizza Monferrato (AT)
Italy

Requirements:
Anforderungen:

Swiss ordinance on measuring instruments (SR
941.210) of February 15, 2006 annex 2 module B;

Directive 2004/22/EC of the European Parliament and
Council of March 31, 2004 on measuring instruments
(MID), appendix B, class of measuring instruments
MI-001.

Conformity of superseded standard:
EN 14154 Edition 2007

Type of instrument:
Geräteart:

**Single-jet impeller meter intended for the metering
of hot water**

Type designation:
Typenbezeichnung:

**DARC, DARC-P, DALC
DBRC, DBRC-P, DBLC, DBLCR2, DBLCH
DPRC, DPRC-P, DPLC, DPLCR2, DPLCH**

Accuracy class(es):
Genauigkeitsklasse(n):

2 (according to OIML R 49:2006)

Characteristics:
Kenndaten:

MAP 16, DN 15 ... DN 40, T70, T90, Class B
Q₃ 1.6 ... Q₃ 16.0

Certificate valid until:
Zertifikat gültig bis:

August 07, 2021

CH-3003 Bern-Wabern, August 08, 2011

Notified body:
Benannte Stelle:

Certificate authority METAS-Cert
No. 1259

For the test:
Für die Prüfung:

Dr. Hugo Bissig

Jürg Ramseyer, Head of METAS-Cert