



Type examination certificate

No. CH-MI001-10037-00

Bauartprüfzertifikat

Applicant:
Auftraggeber:

G.GIOANOLA SRL
Strada Alessandria 50
14049 Nizza Monferrato (AT)
Italy

Requirements:
Anforderungen:

Swiss ordinance on measuring instruments (SR 941.210) of February 15, 2006 annex 2 module B;

Directive 2004/22/EC of the European Parliament and Council of March 31, 2004 on measuring instruments (MID), appendix B, class of measuring instruments MI-001.

Conformity of superseded standard:
EN 14154 Edition 2007

Type of instrument:
Geräteart:

Single-jet impeller meter intended for the metering of cold water

Type designation:
Typenbezeichnung:

**DARF, DARF-P, DALF
DBRF, DBRF-P, DBLF, DBLFR2, DBLFH
DPRF, DPRF-P, DPLF, DPLFR2, DPLFH**

Accuracy class(es):
Genauigkeitsklasse(n):

2 (according to OIML R 49:2006)

Characteristics:
Kenndaten:

MAP 16, DN 15 ... DN 40, T30, T50, Class B
Q₃ 1.6 ... Q₃ 16.0

Certificate valid until:
Zertifikat gültig bis:

June 8, 2020

CH-3003 Bern-Wabern, June 9, 2010

Notified body:
Benannte Stelle:

Certificate authority METAS-Cert
No. 1259

For the test:
Für die Prüfung:

Dr. Hugo Bissig

Jürg Ramseyer, Head of METAS-Cert